

Prothrombin Time (PT)

PT reagent Thromboplastin from rabbit brain origin with a low or high ISI intended for in vitro determination of Prothrombin time according to Quick in citrated human plasma.

BE PT LI Thromboplastin low ISI

- Lyophilized Thromboplastin of rabbit cerebral tissue
- Reconstitution diluent included in the set.
- Ready to use right after reconstitution.
- Low ISI: 1.1 to 1.3
- No interferences with: Triglycerides <731 mg/dL, Bilirubin <228 µmol/L, Hemoglobin <258 µmol/L, Low molecular weight Heparin <0.114 IU, Unfractionated Heparin <0.038 IU



Stability (reconstituted): 5 days at 2 - 8 °C

OBS

72 h at 16 - 18 °C

6 h at 37 °C

Box sizes:

REF 771100: RE 5 x 5 ml; DIL 2 x 15 ml

REF 771101: RE 8 x 12 ml; DIL 8 x 12 ml



BE PT HI Thromboplastin high ISI

- Lyophilized Thromboplastin of rabbit cerebral tissue
- Reconstitution diluent included in the set.
- Ready to use right after reconstitution.
- High ISI: 1.5 to 1.7
- No interferences with: Triglycerides <774 mg/dL, Bilirubin <162 µmol/L, Hemoglobin <258 µmol/L, Low molecular weight Heparin <0.114 IU

Stability (reconstituted): 5 days at 2 - 8 °C

OBS

24 h at 18 - 22 °C

8 h at 37 °C

Box sizes:

REF 771150: RE 5 x 5 ml; DIL 2 x 15 ml

REF 771151: RE 8 x 12 ml; DIL 8 x 12 ml

Activated Partial Thromboplastin Time (APTT)

APTT reagent containing phospholipids Cephalin from rabbit brain origin and contact activator Kaolin or Silica intended for in vitro determination of activated partial thromboplastin time in citrated human plasma.

BE APTT K APTT Kaolin + CaCl₂

- Lyophilized Cephalin (rabbit cerebral tissues) and activator Kaolin
- Calcium chloride solution included in the set.
- Ready to use right after reconstitution.
- The use of Kaolin minimizes the reading time and optimizes the optical detection.
- Insensitive to lupus anticoagulant
- No interferences with: Triglycerides <731 mg/dL, Bilirubin <133 µmol/L, Hemoglobin <261 µmol/L



Stability (reconstituted): 30 days at 2 – 8 °C

OBS

240 h at 18 – 22 °C

Box sizes:

REF 771200: AC 5 x 3 ml; CC 2 x 10 ml

REF 771201: AC 8 x 10 ml; CC 8 x 10 ml

BE APTT SL APTT Silica + CaCl₂

- Lyophilized Cephalin (rabbit cerebral tissues) and activator Silica
- Calcium chloride solution included in the set.
- Ready to use right after reconstitution.
- Insensitive to lupus anticoagulant
- No interferences with: Triglycerides <731 mg/dL, Bilirubin <124 µmol/L, Hemoglobin <261 µmol/L

Stability (reconstituted): 30 days at 2 - 8 °C

OBS

144 h at 16 - 18 °C

Box sizes:

REF 771250: AC 5 x 3 ml; CC 2 x 10 ml

REF 771251: AC 8 x 10 ml; DIL 8 x 10 ml



D-Dimer

D-Dimer reagent is a latex based immunoassay intended for quantitative in vitro determination of D-Dimer in citrated human plasma..

BE DDimer Turbidimetric Immunoassay

- Ready to use latex reagent containing polystyrene particles coated with monoclonal antibodies.
- Dilution buffer and Calibrator are included in the set.
- Wavelengths 400 – 800 nm
- Measuring range 100 – 3200 ng/mL
- Cut-off 200 ng/mL
- Test with a sensitivity of 95 %, specificity of 64 % and negative predictive value (NPV) of 98 %
- No interferences with: Triglycerides <290 mg/dL, Bilirubin <855 µmol/L, Hemoglobin <1,86 mmol/L, Low molecular weight Heparin <100 U/mL, Unfractionated Heparin <100 U/mL

Stability (reconstituted):

8 weeks at 2 – 8 °C

2 weeks at 15 – 25 °C

OBS

120 h at 18 – 22 °C

Box sizes:

REF 771500: BU 3 x 7 ml; AC 3 x 4 ml; CAL 2 x 1 ml; DIL 2 x 7 ml



Coagulation Factor Assays

Deficient plasma of human origin intended for in vitro determination of Factor II, V, VII, X and Factor VIII, IX, XI, XII activity in citrated human plasma.

Deficient plasma

- Lyophilized plasma of human origin free of the specific factor by selective immune adsorption.
- Additional needed diluent BE Owren Buffer.
- PT Factors (Factor II, V, VII, X) no interferences with: Triglycerides <450 mg/dL, Bilirubin <228 µmol/L, Hemoglobin <258 µmol/L, Low molecular weight Heparin <0.114 IU, Unfractionated Heparin <0.038 IU
- APTT Factors (Factor VIII, IX, XI, XII) no interferences with: Triglycerides <731 mg/dL, Bilirubin <124 µmol/L, Hemoglobin <261 µmol/L

Stability (reconstituted): 8 weeks at 2 – 8 °C

OBS

4 h at 18 – 22 °C



BE Factor II Deficient plasma

REF 771602: 6 x 1 ml

BE Factor V Deficient plasma

REF 771605: 6 x 1 ml

BE Factor VII Deficient plasma

REF 771607: 6 x 1 ml

BE Factor VIII Deficient plasma

REF 771608: 6 x 1 ml

BE Factor IX Deficient plasma

REF 771609: 6 x 1 ml

BE Factor X Deficient plasma

REF 771610: 6 x 1 ml

BE Factor XI Deficient plasma

REF 771611: 6 x 1 ml

BE Factor XII Deficient plasma

REF 771612: 6 x 1 ml

Calibrators

Calibration plasma of human origin for quantitative in vitro determination of coagulation tests in citrated human plasma.

BE Cal Set Calibration Plasma PT

- Lyophilized citrated plasma of human origin
- Three different levels of calibration plasma for Prothrombin Time (% INR)
- BE Cal Set values are traceable to WHO INTERNATIONAL STANDARD THROMBOPLASTIN, RABBIT, PLAIN (RBT) from NIBSC



Stability (reconstituted): 4 h at 2 - 25 °C

Box sizes:

REF 775200: CAL1 4 x 0.5 ml; CAL2 4 x 0.5 mL;
CAL3 4 x 0.5 ml



BE Cal Ref Reference Plasma

- Lyophilized citrated plasma of human origin
- Calibration plasma for Fibrinogen and Coagulation Factors
- BE Cal Ref values are determined against its secondary standards of the corresponding primary International Standard for relevant parameters: SSC/ISTH Secondary Coagulation Standard NIBSC code: SSCLOT4

Stability (reconstituted): 4 h at 2 – 25 °C

Box sizes:

REF 775100: CAL 6 x 1 ml

Controls

Control plasma of human origin for quality control of in vitro coagulation tests in citrated human plasma.

BE Trol 1 Plasma Level 1

- Lyophilized citrated plasma of human origin
- Control plasma for Prothrombin Time (PT %, INR, sec.), Activated Partial Thromboplastin Time (APTT sec.), Fibrinogen (mg/dL), Thrombin Time, Factor II (%), Factor V (%), Factor VII (%), Factor VIII (%), Factor IX (%), Factor X (%), Factor XI (%) and Factor XII (%)

Stability (reconstituted):

10 h at 2 – 25 °C

5 days at -20 °C

Box sizes:

REF 773100: CON 6 x 1 ml



BE Trol 2 Plasma Level 2

- Lyophilized citrated plasma of human origin
- Control plasma for Prothrombin Time (PT %, INR, sec.), Activated Partial Thromboplastin Time (APTT sec.), Fibrinogen (mg/dL), Thrombin Time, Factor II (%), Factor V (%), Factor VII (%), Factor VIII (%), Factor IX (%), Factor X (%), Factor XI (%) and Factor XII (%)

Stability (reconstituted):

10 h at 2 – 25 °C

5 days at -20 °C

Box sizes:

REF 773101: CON 6 x 1 ml



BE DDTrol 1 DDimer Plasma low

- Lyophilized plasma of human origin enriched with D-Dimer
- Control plasma for immunoturbidimetric determination of D-Dimer (ng/mL DDU)

Stability (reconstituted):

24 h at 20 – 25 °C

7 days at 2 – 8 °C

Box sizes:

REF 773200: CON 6 x 1 ml



BE DDTrol 2 DDimer Plasma high

- Lyophilized plasma of human origin enriched with D-Dimer
- Control plasma for immunoturbidimetric determination of D-Dimer (ng/mL DDU)

Stability (reconstituted):

24 h at 20 – 25 °C

7 days at 2 – 8 °C

Box sizes:

REF 773201: CON 6 x 1 ml



Cleaning Solution

Cleaning solution intended for cleaning the needle on Behnk automated systems in order to avoid cross contaminations.

BE Clean

- Cleaning solution, acid based
- Ready to use reagent suitable for all Thrombolyzer systems
- Storage at 2-25 °C

Stability (once opened):

OBS

6 days at 15 – 25 °C

Box sizes:

REF 771800: CL 16 x 15 ml



Buffers

Plasma dilution buffer intended for diluting plasma, calibrator and controls during the in vitro determination of coagulation tests.

BE FIB Buffer Fibrinogen Buffer

- Dilution buffer for determination of Fibrinogen on semi-automated systems (plasma dilutions with 450 µl buffer on Thrombostat or 950 µl buffer on Thrombotimer).
- Same buffer as enclosed in the BE FIB set.

Stability (once opened):

OBS

6 days at 15 – 25 °C

Box sizes:

REF 771350: BU 16 x 15 ml

BE Owren Buffer Owren Koller Buffer

- Dilution buffer for determination of Factor II, Factor V, Factor VII, Factor VIII, Factor IX, Factor X, Factor XI and Factor XII

Stability (once opened):

OBS

6 days at 15 – 25 °C

Box sizes:

REF 771700: BU 16 x 15 ml

